



QPS Evaluation Services Inc
Testing, Certification and Field Evaluation Body
Accredited in Canada, the USA, and Internationally

File
LR1356

CERTIFICATE OF COMPLIANCE
(ISO TYPE 3 CERTIFICATION SYSTEM)

Issued to	SVANTEK Sp. z o.o
Address	Strzyglowska 81 04-872 Warszawa Poland
Project Number	LR1356-1
Product	Intrinsically safe personal noise dosimeter
Model Number	SV 104IS
Ratings	Single lithium-ion cell Um=6.6V Ex ia IIC T4 Ga Class I, Zone 0, AEx ia IIC T4 Ga T _{Amb} = -10°C to +50°C
Applicable Standards	CAN/CSA C22.2 No 61010-1 CAN/CSA C22.2 No 60079-0 CAN/CSA C22.2 No 60079-11 ANSI/UL 61010-1 ANSI/UL 60079-0 ANSI/UL 60079-11
Factory/Manufacturing Location	Same as Applicant

Statement of Compliance: The product(s) identified in this Certificate and described in the Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.



Issued By: **Dave Adams, P. Eng.**
Manager, Hazardous Locations Dept. [Ex Equipment]

Signature:

Date: October 22, 2015